

Fitbone™

Intramedullary Lengthening System
Surgical Technique

Fitbone Transport and Lengthening System Tibia Application

(Bi-cortical Screw Fixation Technique)





Table Of Contents

4	Surgery planning - tibia
4	Fitbone Transport and Lengthening System nail types
6	Surgery
8	Insertion of the Half Pins (Selfdrilling XCaliber Cylindrical Screw Shaft) for torsion control
10	Approach and initial reaming
13	Reaming of the medullary cavity
13	Reaming of the medullary cavity with rigid reamers
16	Initial reaming
16	Reaming of the medullary cavity
19	Extracting the working tube
19	Reaming of the medullary cavity with flexible reamers
21	Dummy TAA/Trial nail TN/TLN and Jig assembly
22	Insertion of the Dummy TAA/Trial Nail TN/TLN
23	Potential problems and solutions
25	Blocking screws insertion
26	Removal of the recharging screw
27	Insertion of the Fitbone TN/TLN
35	Fibula osteotomy (only in bone transport + lengthening cases)
36	Final intraoperative test
38	Docking site compression and healing
38	Lengthening (optional) with the Fitbone TN/TLN
41	Fitbone Bone Transport and Lengthening System – removal opt – tibia
50	Damaged screw removal

The surgical technique shown is for illustrative purposes only. The technique(s) actually employed in each case will always depend upon the medical judgment of the surgeon exercised before and during surgery as to the best mode of treatment for each patient. Please kindly refer to the product IFU PQFBT and to the reusable medical devices IFU PQRMD that contain instructions for use of the product.

SURGERY PLANNING - TIBIA

The following points are considered during pre-operative planning:

- Deformity correction (if required)
- Defect size and location
- Medullary canal diameter
- Optimal nail length
- Location of bone transport screw (maintain at least 10mm between the transport screw and resection site/osteotomy site)
- Location of blocking screws (maintain at least 10mm, or more in case of poor bone quality, between the blocking screw and resection site/osteotomy site)
- Location of osteotomy (maintain at least 10mm between screw (locking or blocking) and osteotomy)
- Required additional lengthening (for TLN)
- Required nail "rewind and go" (see page 32) and new location of the bone transport screw by recharging the nail.



Fig. 1 Fitbone Transport and Lengthening System - Tibial Nail

TN and TLN trial nails are optionally available and should be ordered separately.

Please refer to the next point for all available Fitbone TN and TLN lengths diameters and bone transport / lengthening strokes available.

FITBONE TRANSPORT AND LENGTHENING SYSTEM NAIL TYPES

Fitbone Transport Nails (Fitbone TN)

Fitbone TN Tibia




Fitbone Transport and Lengthening Nails (Fitbone TLN)

Fitbone TLN Tibia




Standard Locking Screws



Ø	L
4.5mm	From 20 to 80 (2.5mm increment)

Revision Locking screws



Ø	L
4.5mm	From 20 to 80 (2.5mm increment)

SURGERY

Grid Plate

Part #	Description
60001464	FITBONE GRID PLATE

The GRID plate is recommended in FITBONE surgeries to ensure the correct limb alignment and deformity correction (if needed). The GRID plate is placed on the operating table underneath the cushion. To prevent damage to the GRID plate, it should rest flat over its entire surface.

Never place the GRID plate directly under the patient without padding.

The Grid plate has two double lines. The joint centres are placed over these double lines to ensure correct alignment.

Technical features of the grid

Material: Pertinax RI4 0000 (Hartpapier) PF CP 201 HP 2061

Dimensions: 37.6cm x 128.2cm

50mm between longitudinal and transverse wires.

1mm between double wires (**shown in bold in Fig.2**).

Patient positioning

The patient is placed supine on the operating table. The operated leg is draped to allow free movement. Ensure that the C-arm can produce an uninterrupted view of the hip, knee and ankle joint.

Ensure that the C-arm is perpendicular to the GRID plate and that the joints are viewed in the centre of the screen. This will minimize parallax errors.

Place the centre of the hip joint over the double radiopaque grid line while ensuring the patella faces forward.

Place the ankle joint over the double radiopaque grid line without changing limb rotation. (**Fig. 3**)

The center of the knee joint might not be on the double radiopaque grid line if a deformity is present. (**Fig. 4**)

Proceed with next steps when the alignment matches the pre-operative plan.




Fig. 2 GRID Plate




Fig.3 Patient positioning




Fig. 4 Hip, knee and ankle joint position

Patient marking

Part #	Description
99-6000xxxx	FITBONE TRANSPORT TRIAL NAIL
99-6000xxxx	FITBONE TRANSPORT LENGTHENING TRIAL NAIL
60003516	FITBONE TRANSPORT SIZING GUIDE
60003517	FITBONE TLN SCREW OVERLAY
60003518	FITBONE TIBIA/FEMUR RETRO TN SCREW OVERLAY

With the image intensifier, locate the knee joint line using a wire placed over the skin (**Fig. 5**).

The nail marking can be done by using the sizing guide or the correspondent Fitbone TN/TLN trial nail if available.

The sizing guide allows to check for diameter, total length, slot's length and position of the implant. (**Fig. 5a, 5b, 5c**) When using the sizing guide place the corresponding TN/TLN screw overlay on the holes of the sizing guide at the correct length to check the position of the distal holes (**Fig. 5d**).

In case of using a trial nail, this should be placed deep into the entry point..




Fig. 5 Knee joint marking




Fig.5a Sizing guide and Screw Overlay




Fig. 5b Sizing guide distal features




Fig. 5c Sizing guide slot length and position




Fig. 5d Sizing guide proximal features

Mark the end of the nail and the osteotomy level according to the pre-operative plan change in diameter of the nail with a ruler and the pen and in addition with a surgical stapler (**Fig. 6**). Optionally, the nail contour and the screws position can also be marked according to the dimensions of the nail to be used.




Fig. 6 Osteotomy and nail position marking

INSERTION OF THE HALF PINS (SELFDRILLING XCALIBER CYLINDRICAL SCREW SHAFT) FOR TORSION CONTROL

Part #	Description
99-941550	SELFDRILLING XCALIBER CYLINDRICAL SCREW SHAFT D6MM THREAD 5MM L180/50QC STERILE



PRECAUTION: Half-pins might be used to avoid unintentional axial and torsional deviations. These must be placed in such a way that they do not interfere with the insertion of the intramedullary nail into the medullary canal.

With the leg in full extension, insert one 4.5 or 5mm half-pin, to ensure correct torsional alignment.

Insert the half-pin in the supracondylar region, posterior to the path of the nail, in the coronal plane lateral to medial or in the posterior part of the proximal tibia, at surgeon's discretion. (**Fig. 7**)




Fig. 7 Distal femur half-pin position

The half-pin should be positioned perpendicular to the foot if torsional corrections are not needed. **(Fig. 8)**

The position of the half-pin must be adapted based on bone defect position.

The half-pin is placed as a visual cue to ensure rotational alignment. At the end of the procedure, the half-pin should be perpendicular to the foot. **(Fig. 8)**




Fig. 8 Final position half-pin and foot

Venting holes

Venting holes should be drilled at the osteotomy level to reduce pressure during reaming and allow bone debris to exit. **(Fig. 9)**



Fig. 9 Venting holes

PATIENT POSITIONING

Infrapatellar approach

Place the knee in maximum flexion during the surgical procedure. **(Fig. 10)**




Fig. 10 Patient positioning infrapatellar approach

Suprapatellar approach

A sufficiently sized sterile covered removable support (**Fig. 11**) should be used to maintain the knee flexed at about 30° during the surgical procedure.




Fig. 11 Patient positioning suprapatellar approach

APPROACH AND INITIAL REAMING

Infrapatellar approach

Make a 20mm transverse skin incision between the lower border of the tibial patella and the tibial tuberosity (**Fig. 12**).

Optionally, it is possible to perform a longitudinal incision centered over the distal pole of the patella.




Fig. 12 Skin incision infrapatellar approach

Split the patella tendon longitudinally (**Fig. 13**).




Fig. 13 Patella tendon incision infrapatellar approach

Suprapatellar approach

Make a 20mm transverse skin incision 20mm above the patella (**Fig. 14**).




Fig. 14 Skin incision suprapatellar approach

Split the quadriceps tendon longitudinal (**Fig. 15**).




Fig. 15 Patella tendon incision suprapatellar approach

Kirschner wire entry points

Part #	Description
60001039	FITBONE KIRSCHNER WIRE D3 L280

Infrapatellar approach

Locate the center of the bone canal under x-ray guidance using an AP-view. Insert a kirschner wire at the bony tibial plateau, through the patella tendon, with the knee in flexion (**Fig. 16**). The kirschner wire should not be inserted more than 2mm away from the target position. Direct the kirschner wire according to your planning, directing the wire along the intended ultimate nail path (in case deformity correction is required) and parallel to the anterior cortex of the tibia. The kirschner wire can be inserted to the osteotomy level to guide the rigid reamer.




Fig. 16 Kirschner wire insertion at maximum knee flexion (infrapatellar approach)

Double-check position with the C-arm AP and lateral (Fig. 17).




Fig. 17 Kirschner wire position (infrapatellar approach)

Suprapatellar approach

With the knee in 30° flexion, insert a 3mm kirschner wire at the anterior aspect of the tibia, through the quadriceps tendon, (Fig. 18) behind the patella. The kirschner wire should be inserted exactly at the target position. Direct the kirschner wire according to your planning, directing the wire along the intended ultimate nail path (in case deformity correction is required) and parallel to the anterior cortex of the tibia.




Fig. 18 Kirschner wire insertion (suprapatellar approach)

Double-check position with the C-arm AP and lateral (Fig. 19).




Fig. 19 Kirschner wire position (suprapatellar approach)

REAMING OF THE MEDULLARY CAVITY

Fitbone Transport and Lengthening system instruments include rigid and flexible reamers, to be used as surgeon's discretion.

Next section will describe the reaming procedure with rigid reamers and the following section the reaming procedure with flexible reamers".

REAMING OF THE MEDULLARY CAVITY WITH RIGID REAMERS

The use of 480mm long rigid reamers is limited for nails up to 350mm long.

Cones and tubes

Part #	Description
60001014	Fitbone™ tube T14/13-M
60001015	Fitbone™ tube T13/12-M
60001016	Fitbone™ tube T12/11-M
60001017	Fitbone™ tube T12/10-M
60001018	Fitbone™ tube T12/09-M
60001054	Fitbone™ tube T14/13-L
60001055	Fitbone™ tube T13/12-L
60001020	Fitbone™ tube T12/11-L
60001021	Fitbone™ tube T12/10-L
60001022	Fitbone™ tube T12/09-L
60001054	Fitbone™ tube T14/13-XL
60001059	Fitbone™ tube T13/12-XL
60001024	Fitbone™ tube T12/11-XL
60001025	Fitbone™ tube T12/10-XL
60001026	Fitbone™ tube T12/09-XL
60001888	Fitbone™ Cone C13 Retropatellar
60001028	Fitbone™ cone C13
60001029	Fitbone™ cone C13+
60001030	Fitbone™ cone C13++
60001033	Fitbone™ tube-sinker TS13

Cones and tubes are recommended to perform minimally invasive accurate reaming.

For 11mm diameter Fitbone TN/TLN nails, 13mm cones are required

Infrapatellar approach

With a perfect kirschner wire placement, insert the centered cone C13 (external diameter 13mm). If the position of the kirschner wire has to be corrected, use excentric C13+, or C13++ to correct 1 or 2mm in any direction. The notch at the end of the cone will help in achieving the correct orientation. If correction of more than 2mm is needed, remove and re-insert the kirschner wire in the corrected position.

Optionally, before inserting the cone over the kirschner wire, slide the tube T14/13-S or -M ("Working Tube") over the cone (Fig. 20).




Fig. 20 Cone variations




Fig. 21 Infrapatellar cone




Fig. 22 Suprapatellar Cone insertion



Fig. 23 Suprapatellar Cone

Suprapatellar approach

When using the suprapatellar approach, use the Cone C13RP (CONE C13 RETROPATELLAR). The conical part of the C13 RP suprapatellar cone is 20mm longer than the standard approach cones and has a smooth conical surface to aid insertion (Fig. 22 and 23). Before insertion slide the tube T14/13-S or -M ("Working Tube") over the cone. After passing the patella use the tube-sinker TS13 and push the tube just in front of the tibia plateau. Then replace the C13 RP by an excentric cone C13+, or C13++ to correct 1 or 2mm in any direction.

Cones insertion

Infrapatellar

Using a mallet and the cone sinker (CS 15-13), insert the cone into the bone. To ensure stability the cone should be inserted at least up to the depth of the teeth. (Fig. 24)

Suprapatellar

Using a mallet and the cone sinker (CS 15-13), insert the cone C13RP into the bone.

If the bone size and stability allow, insert the conical part of the tube into the bone. This will prevent damaging the tube. If the suprapatellar cone is not available use a normal cone and insert it at least up to the depth of the teeth.



PRECAUTION: Do not directly hammer the cone because that will damage the cone end and tubes will not slide on the cone as required.



Fig. 24 Cone insertion

REF	Cone Sinker	REF	Cone	REF	Working Tube
60001036	CS 15-13	60001028	C13	60001014	Tube T14/13
		60001029	C13+		
		60001030	C13++		
60001036	CS 15-13	60001888	C13RP	60001014	Tube T14/13

Table 1 Cones, Cone Sinker and Working tube assembly

Tubes

The Fitbone Tubing System has several key functions to assist in delivering a successful surgery. The tubes protect soft tissues, guide the reamers along the planned alignment, assist in the removal of reamed bone debris and protect the bone canal once reamed.

In order to guide the reamers along the planned alignment, tubes are placed in sequence inside one another. **Table 2** provides an overview of which tubes should be used with each reamer.

Please note that at least 0.5mm of space is needed between the reamer and the inner diameter of a tube to prevent material from blocking the reamer while reaming.

There are four different tube lengths S (Small), M (Medium), L (Large) and XL (Extra Large). (Fig. 25)

Depending on the Set provided, not all tubes may be available. There is the possibility to ream with flexible reamers instead of rigid reamers in cases of no deformity correction or by surgeons' decision.




Fig. 25 Tubes (S, M, L and XL)

Tube	S REF	M REF	L REF	XL REF	Rigid Reamer	Front cutting Rigid Reamer REF	Rounded Rigid Reamer REF
Tube T12/09	60001048	60001018	60001022	60001026	Reamer D8.0		60000411
Tube T12/10	60001047	60001017	60001021	60001025	Reamer D9.0	60000412	60000413
Tube T12/11	60001046	60001016	60001020	60001024	Reamer D10.0	60000414	60000415
Tube T13/12	60001045	60001015	60001055	60001059	Reamer D10.5		60000416
					Reamer D11.0	60000417	60000418
					Reamer D11.5		60000833
Tube T14/13	60001044	60001014	60001054	60001058	Reamer D12.0	60000419	60000420

Table 2 Tubes and Rigid Reamers assembly

Working tube insertion

Use the tube-sinker TS15 to insert the tube into the bone, according to **Table 3**.

For Fitbone TN/TLN 13, Tube14/13 must be used (short or medium as available)

Tube	S REF	M REF	L REF	XL REF	Tube Sinker	REF
Tube T12/09	60001048	60001018	60001022	60001026	TS 13	60001033
Tube T12/10	60001047	60001017	60001021	60001025		
Tube T12/11	60001046	60001016	60001020	60001024		
Tube T13/12	60001045	60001015	60001055	60001059		
Tube T14/13	60001044	60001014	60001054	60001058		

Table 3 Tubes and Tube sinker assembly



Fig. 26a Working tube insertion

The working tube should be inserted about 5-10mm into the bone, making sure the tube is stable inside the bone and the insertion angle of the reamer is orientated according to the anticipated direction of the nail based on preoperative plan. For the infrapatellar approach use either an S or M T14/13 tube, and an M or L for the suprapatellar approach. **(Figure 26a and 26b)**

Remove the cone and kirschner wire, leaving the working tube firmly in place.



WARNING: If an eccentric cone has been used, do not rotate it while removing.




Fig. 26b Final working tube position

INITIAL REAMING

Part #	Description
60000411	Fitbone™ reamer D8.0 L200 L480 rounded
60000413	Fitbone™ reamer D9.0 L200 L480 rounded
60000412	Fitbone™ reamer D9.0 L100 L480 forehead cutting
60000415	Fitbone™ reamer D10.0 L200 L480 rounded
60000414	Fitbone™ reamer D10.0 L100 L480 forehead cutting
60000416	Fitbone™ reamer D10.5 L200 L480 rounded
60000418	Fitbone™ reamer D11.0 L200 L480 rounded
60000417	Fitbone™ reamer D11.0 L100 L480 forehead cutting
60000833	Fitbone™ reamer D11.5 L200 L480 rounded
60000420	Fitbone™ reamer D12.0 L200 L480 rounded
60000419	Fitbone™ reamer D12.0 L100 L480 forehead cutting
60000716	Fitbone™ reamer D12.5 L200 L480 rounded
60000392	Fitbone™ T-handle
60001179	Fitbone™ step reamer TAA1180
60001415	Fitbone™ step reamer TAA1160
60001528	Fitbone™ step reamer TAA1140

The Fitbone System includes two different types of reamers (**Fig. 27**):

1. Rounded reamer (Fitbone reamer rounded)
2. Front cutting reamer (Fitbone reamer forehead cutting)

The rounded rigid reamers have a cutting length of 200mm and are used to open and straighten the medullary cavity. To secure the entry point at all times and to avoid unintended displacements, the use of tubes is strongly recommended.

The front cutting rigid reamers with a cutting length of 100mm make it possible to open and correct medullary cavity reaming in any direction.



WARNING: There is a risk of excessive cortical weakening and even perforation, leading to a fracture during treatment, particularly with the front cutting reamers.

REAMING OF THE MEDULLARY CAVITY

REAMING STEPS with rigid reamers

- Ream up to the osteotomy
- Perform osteotomy
- Ream to the end of the nail
- Perform final ream with step reamer

These steps allow acute correction according to preoperative-plan.

It is very important to follow the reaming steps mentioned above in order to avoid breaching the posterior cortex.

Insert the 12mm front cutting rigid reamer through the working tube and ream the bone to a depth of 2cm.




Fig. 27 Front cutting reamer (Left), rounded reamer (Right)




Fig. 28 Reducing the reaming channel

Reduce tube diameter for concentric reaming by inserting the tubes T13/12 and T12/09 or T12/10 in case of 11mm nails (**Fig. 28**). Ream with an 8mm or 9mm rounded rigid reamer up to the planned osteotomy level, indicated by the skin marker, and assisted by AP and lateral fluoroscopy.



PRECAUTION: Monitor the entire reaming process with the image intensifier in two planes to detect any reaming errors in time.

Reaming of proximal segment

The proximal segment corresponds to the proximal section of the nail, with a larger diameter. Ream the proximal segment as follows (**Fig. 29**):

Continue with 10mm, 11mm and finally with 12mm rigid reamer, changing the inner tubes corresponding to the rigid reamer in use, see also **Table 2**.

There is also a 12.5mm rigid reamer available if required.

Double-check the canal after each reaming with AP and lateral fluoroscopy.

The inner tube diameter is the smallest number and the outer diameter the largest, e.g. 13/12 means 13mm outer and 12mm inner diameter.



Fig. 29 Reaming up to the osteotomy

Osteotomy

Perform a percutaneous osteotomy at the previously marked level (**Fig. 30**). Ensure that the osteotomy is completed before continuing.

Once the osteotomy has been performed, manipulate the limb into the correct alignment according to the preoperative plan (Reverse Planning Method).

Ensure correct rotational alignment of the limb by keeping the reference half-pin perpendicular to the foot.

An external fixator can be attached to the half-pins to retain the corrected alignment.

Upon completion of osteotomy and bone correction, begin reaming of the distal fragment. Following the same process as before, reaming the distal segment to the planned tip of the nail, as indicated by the skin marker.




Fig. 30 Osteotomy finalization

Reaming of distal segment

The distal segment corresponds to the distal section of the nail, with a smaller diameter. Ream the distal segment as follows:

Gradually ream from 8mm or 9mm up to 11mm using rigid rounded reamers.



PRECAUTION: Do not weaken the cortex as this could increase the risk of fracture during treatment (Fig. 31).

If the correct position cannot easily be reached, use the front cutting reamers to carefully widen the midshaft canal where it is needed. The front cutting reamer will aggressively remove part of the sclerotic or cortical bone, thus the process has to be controlled under AP and lateral fluoroscopy at all times.



WARNING: Never use front cutting rigid reamers in the area of the final position of the tip of the Fitbone intramedullary nail.

Step reamer TAA

The final ream is performed using the appropriate step reamer (Fig. 32).

Perform reaming by hand, using the T-handle in order to achieve better control.

The canal should be reamed up to the point where the change in diameter of the step reamer correspond to where the change in diameter of the implant is going to be, which should be already marked on the skin. (Fig. 32a and 32b)"

For every implant size, there is a step reamer TAA that can be used (see table 4).

The step reamer should be used with the working tube T14/13.




Fig. 31 Reaming with rigid reamers



Fig. 32 Reaming of distal segment and step reamer




Fig. 32a Step reamer TAA final position

Fig. 32b Level of the change in diameter of the nail

Fitbone TN/TLN	Step Reamer TAA	REF
Fitbone TRANSPORT NAIL TN 1140-T-XXX Fitbone TRANSPORT LENGTHENING NAIL TLN 1140-T-XX-XXX	Step Reamer TAA1140	60001528
Fitbone TRANSPORT NAIL TN 1160-T-XXX Fitbone TRANSPORT LENGTHENING NAIL TLN 1160-T-XX-XXX	Step Reamer TAA1160	60001415
Fitbone TRANSPORT NAIL TN 1180-T-XXX Fitbone TRANSPORT LENGTHENING NAIL TLN 1180-T-XX-XXX	Step Reamer TAA1180	60001179

Table 4 - Overview Fitbone TN/TLN and Step Reamer TAA

EXTRACTING THE WORKING TUBE

Part #	Description
60001038	Fitbone™ clamp

The working tube is removed using the tube extractor clamp (**Fig. 33**).




Fig. 33 Extracting the working tube

REAMING OF THE MEDULLARY CAVITY WITH FLEXIBLE REAMERS

Part #	Description
172991	Flexible Reamer System Box
99-173281	Guide Wire with Olive Ø3x980mm
173230	reamer sleeve
183260	Awl
60001039	FITBONE KIRSCHNER WIRE D3 L280

Insert the kirschner wire into the intramedullary canal, via the piriformis fossa using fluoroscopy in AP and lateral view .

Insert the tissue protector (reamer sleeve) over the kirschner wire at the level of the entry point followed by the awl.

Open the entry point by gently advancing the awl over the kirschner wire with a twisting motion up to the marked part. (**Fig. 34**)

Remove the awl, the kirschner wire and the tissue protector (reamer sleeve).

In case the reamer sleeve is too big to be used in the supra- or infra-patellar approach, the Fitbone TAA cone and tube system can be used as mentioned in the section "Reaming of the medullary cavity with rigid reamers". In this case use the reamer 12mm or 12.5mm as entry reamer

A 3x980mm guide wire with olive is inserted into the medullary canal, and its position confirmed in both planes.

Proceed with reaming of the medullary canal.

Ream with a 8mm or 9mm flexible reamer up to the planned end of the nail, indicated by the skin marker, and assisted by AP and lateral fluoroscopy. Increase reaming gradually of 0.5mm increments, until the intramedullary channel is overreamed by 2mm. (**Fig. 35**)




Fig. 34 Awl



Fig. 35 Reaming of intramedullary channel



PRECAUTION: Monitor the entire reaming process with the image intensifier in two planes to detect any reaming errors in time.

Reaming of proximal segment

The proximal segment corresponds to the proximal section of the nail, with a larger diameter (12mm), which is marked on the skin with a stapler. Ream the proximal segment as follows:

Increase reaming gradually of 0.5mm increments, until the intramedullary channel is overreamed by 2mm according to the size of the nail, up to 14mm.

Double-check the canal after each reaming with AP and lateral fluoroscopy.



PRECAUTION: Do not weaken the cortex as this could increase the risk of fracture during treatment.

OSTEOTOMY

Perform a percutaneous osteotomy at the previously marked level. **(Fig. 30)**

Once the osteotomy has been performed, manipulate the limb into the correct alignment according to the preoperative plan.






Ensure correct rotational alignment of the limb by keeping the reference half-pin perpendicular to the foot.

An external fixator can be attached to the half-pins to retain the corrected alignment.

Upon completion of osteotomy and bone correction, begin reaming of the distal fragment. Following the same process as before, reaming the distal segment to the planned tip of the nail, as indicated by the skin marker.

DUMMY TAA/TRIAL NAIL TN/TLN AND JIG ASSEMBLY

Part #	Description
99-6000xxxx	FITBONE TRANSPORT TRIAL NAIL
99-6000xxxx	FITBONE TRANSPORT LENGTHENING TRIAL NAIL
60000832	Fitbone™ dummy TAA1160-T-225
60001495	Fitbone™ dummy TAA1180-T-245
60001244	Fitbone™ dummy TAA1140-T-205
60000689	Fitbone™ open end wrench SW14/17
60000688	Fitbone™ test pin D4.5
60001184	Fitbone™ outrigger TAA
60000175	Fitbone™ fastening bolt M6 L9
60001175	Fitbone™ drill guide TAA
60000218	Fitbone™ space holder TAA
60000310	Fitbone™ connection bolt TAA
60000219	Fitbone™ clamping nut TAA
60000003	Fitbone™ setscrew

Fitbone Drill Guide TAA	Fitbone Outrigger TAA	Fitbone Space Holder TAA	Fitbone Clamping Nut TAA	Fitbone Connection Bolt TAA
				
The Fitbone drill guide TAA is bilateral since it can be used for left leg "L" or right leg "R"	The Fitbone outrigger TAA is the main body of jig	The Fitbone space holder is the part of the jig to which the Fitbone dummy/nail will be assembled and it has a notch that is used as a reference mark to correctly position the Fitbone into the bone.	The Fitbone clamping nut is used to fix the space holder into the outrigger TAA.	The Fitbone connection bolt TAA is used to fix the dummy/nail through its thread

If a check of the proper assembly of the jig is preferred, the drill guide can be attached and the check can be performed.

Attach the drill guide to the outrigger, paying attention to the indication "L" and "R" (**Fig. 36**) and lock it in place. (**Fig. 37**)




Fig. 36 Marking for right femur "R" or left femur "L"




Fig. 37 Drill guide fixation

Assembly space holder

Insert the Space Holder into the squared hole in the back of the handle, making sure that the marking “THIS SIDE UP” is facing upwards, and lock it by hand using the clamping nut TAA. **(Fig. 38)**




Fig. 38 Space holder correct assembly

Insert the “connection bolt” into the “space holder” and fix the assembly by screwing the “clamping nut” on the thread of the “space holder”. **(Fig. 39)**

The space holder is adapted to both intra- and suprapatellar approach.




Fig. 39 Space holder fixation

INSERTION OF THE DUMMY TAA/TRIAL NAIL TN/TLN

The trial nail TN/TLN is used to confirm the medullary canal is appropriately prepared for the Fitbone nail to be implanted. It must be possible to insert the nail without resistance and without hammering.

The trial nail TN/TLN should be connected to the jig **(Fig. 40)** and inserted to confirm the final nail can be implanted at the planned depth (5-10mm below bone entry point) and alignment.

The dummy nail TAA can be used to place blocking screws if required. It can also be useful to confirm the fitting of the proximal part of the TN/TLN nail into the bone since the geometry of TAA and TN/TLN at the proximal part is the same.

Refer to Table 4 for the correct combination of dummy TAA and step reamer TAA, based on the selected FITBONE TN/TLN.




Fig. 40 Trial nail TN/TLN nail assembly

Trial nail TN/TLN insertion

Insert the trial nail TN/TLN to confirm planned nail position is achieved. The drill guide portion of the jig can be removed for better visualization under x-ray. Ensure the space holder marker on the jig is at the level of the entry point. This indicates that the nail is sufficiently implanted in the bone, to a depth of 5-10mm. It is advised to check nail positioning under fluoroscopy in AP and lateral planes.

(Fig. 41)




Fig. 41 Dummy/Trial nail correctly positioned

POTENTIAL PROBLEMS AND SOLUTIONS

1. The marking "THIS SIDE UP" on the Space Holder is not facing upwards and the connection between drill guide assembly and nail is loose. (Fig. 42)

Corrective action: Place the Space Holder correctly with "THIS SIDE UP" facing upwards.




Fig. 42 Space holder wrongly positioned

2. The Space Holder is not completely flush with the Drill Guide Assembly (Fig. 43).

Corrective action: Insert the Space Holder correctly, making sure that the squared end is placed into the square hole in the back of the handle.




Fig. 43 Space holder wrongly inserted

3. The dummy/trial nail is wrongly placed. (Fig. 44)

Corrective action: Unlock the “Connection Bolt” and rotate the nail 180°.




Fig. 44 Dummy/Trial nail wrongly positioned

4. The dummy/trial nail is loosely fixed to the “connection bolt”. (Fig. 45)

Corrective action: Tighten the Connection Bolt firmly. The fixation has to be tight.




Fig. 45 Connection Bolt loose

BLOCKING SCREWS INSERTION

Part #	Description
700213	DISTAL TISSUE GUIDE
99-700183	DISTAL DRILL BIT D4.5MM QC STERILE
700277	DISTAL SCREW SIZING GAUGE
700320	SHORT SCREWDRIVER
700319	SHORT SCREWDRIVER RETENTION ROD

The use of blocking screws to achieve correct alignment is recommended. The anterior-posterior screw (to the nail) achieves correct alignment in the lateral view while the medial-lateral screw maintains alignment in the AP view (to the nail) (Fig. 46). If blocking screws are needed, required, standard 4.5mm Fitbone screws can be inserted. Please see page 30 for guidance - Inserting free-hand locking screw.

In order to place the blocking screws use the Fitbone TN or TLN trial nail. Alternatively the TAA dummy nail that most resembles the chosen Fitbone TN or TNL nail can be used. See **table 5**.

When using a trial TN/TLN make sure that the distal and proximal fragments are aligned correctly and final stabilization of the Fitbone TN/TLN is ensured.



PRECAUTION: Axial correction is determined definitely by the choice of the entry point for the FITBONE® TAA intramedullary lengthening nail in the bone and the reaming direction up to the osteotomy site.




Fig. 46 Alignment and blocking screws

Fitbone TN	Fitbone TLN	Step Reamer TAA	REF	Dummy TAA	REF
Fitbone TN 1140-T-XXX	Fitbone TLN 1140-T-XX-XXX	Step Reamer TAA1140	60001528	Dummy TAA1140-T-205	60001244
Fitbone TN 1160-T-XXX	Fitbone TLN 1160-T-XX-XXX	Step Reamer TAA1160	60001415	Dummy TAA1160-T-225	60000832
Fitbone TN 1180-T-XXX	Fitbone TLN 1180-T-XX-XXX	Step Reamer TAA1180	60001179	Dummy TAA 1180-T-245	60001495

Table 5. Overview Fitbone TN/TLN, Step Reamer TAA and Dummy TAA

Alignment Control

Remove the jig, leaving the trial nail in place. Place the leg in full extension and control alignment with the double line grid as described previously. (Fig. 47)

The double line should be centred on the hip, and ankle joint, passing the knee joint as planned (RPM). Perform additional reaming or correction alignment if needed.




Fig. 47 Alignment control

REMOVAL OF THE RECHARGING SCREW

Part #	Description
99-10013	Allen Wrench 3mm Sterile (Single Use)

In case the whole stroke of the nail will be needed during the treatment and no retraction of the nail is expected, the recharging screw should be removed by using the Allen Wrench.




Fig. 48 Recharging screw




Fig. 49 Available stroke with recharging screw assembled




Fig. 50 Available stroke without recharging screw

INSERTION OF THE FITBONE TN/TLN

Part #	Description
60000689	Fitbone™ open end wrench SW14/17
60000688	Fitbone™ test pin D4.5
60001184	Fitbone™ outrigger TAA
60000175	Fitbone™ fastening bolt M6 L9
60001175	Fitbone™ drill guide TAA
60000218	Fitbone™ space holder TAA
60000310	Fitbone™ connection bolt TAA
60000219	Fitbone™ clamping nut TAA
60000003	Fitbone™ setscrew

Insert the cable of the Fitbone nail carefully into the connection bolt and attach the Fitbone nail to the Jig (**Fig. 51**). Attach the drill guide to the outrigger and insert the test pins D4.5 into the drill guide and make sure that they pass through the holes in the nail easily without friction. Refer to page 16 for potential problems and solutions.

Lock the nail firmly in place using two open end wrenches SW14/17 simultaneously (**Fig. 52**). Remove the test pins.




Fig. 51 Final nail assembly




Fig. 52 Final nail assembly

Insert the nail into the tibia, making sure that the marking notch is at the level of the entry point (**Fig. 53**) or 5mm outside the bone, so that the nail is inserted into the bone deep enough.



WARNING: Never use a hammer to drive or remove the Fitbone Nail into/from the medullary cavity since this could damage the implant.




Fig. 53 Nail insertion

Inserting the jig assisted locking screws

Part #	Description
700302	PROXIMAL TISSUE GUIDE
700212	TROCAR D4.5MM
700321	LONG SCREWDRIVER
700282	LONG SCREWDRIVER RETENTION ROD
99-700182	PROXIMAL DRILL BIT D4.5MM QC STERILE



PRECAUTION: Please make sure to use the specific instrumentation according to the locking screw model as described in the Operative Technique.

The jig assisted screws have a diameter of 4.5mm. They are available in standard and revision options (**Fig. 54**).

Insert the trocar into the tissue guide. Advance the tissue guide and trocar through the jig to determine the skin insertion point.

Make a 1 cm incision at the insertion point and bluntly dissect through the soft tissues to allow the tissue guide and trocar to pass down to bone.



PRECAUTION: To ensure the correct locking screw length, the tip of the trocar must go flat against the bone.

Secure the tissue guide to the jig by closing the locking cam. Insert the drill through the trocar.

Drill bi-cortically, ensuring the distal tip of the proximal drill penetrates the far cortex.

Do not drill beyond the far cortex. Penetrate the far cortex only.





Fig. 54 Tissue guide and trocar insertion



Screw profile

Measure the screw length using the gauge on the shaft of the drill.



PRECAUTION: Choosing the appropriate length for the locking screw is critical for the effective and safe stabilization of bone fragments. The length of the screw should be determined by using the dedicated instruments as described in the operative technique.

Remove the trocar and the drill leaving the tissue guide in place.



PRECAUTION: After locking screw length measurements, make sure to hold the targeting handle in the same position: any movement of the targeting handle may result a mismatch between the screw and the hole.

Insert the locking screw using the long screw driver with retention rod through the proximal tissue guide and advance until the screw head is seated on the bone (**Fig. 55**).



PRECAUTION: Overtightening of the screw should be avoided and the head of the screw should come just in contact with the cortex. Stop insertion when resistance is felt.

Verify the locking screw position under fluoroscopy.

In instances of poor bone quality and subsequent insufficient fixation when using standard screws, the revision screw option can be utilized.

Remove the tissue guide from the jig by removing the locking nut.

Using the same process, insert the second locking screw.




Fig. 55 Locking screw hole drill, length measurement and screw positioning

Inserting free-hand locking screw

Part #	Description
700213	DISTAL TISSUE GUIDE
99-700183	DISTAL DRILL BIT D4.5MM QC STERILE
99-700184	DISTAL DRILL BIT D4.0MM QC STERILE
700277	DISTAL SCREW SIZING GAUGE
700320	SHORT SCREWDRIVER
700319	SHORT SCREWDRIVER RETENTION ROD



PRECAUTION: Please make sure to use the specific instrumentation according to the locking screw model as described in the Operative Technique.

Before locking screw fixation check that the half-pin is perpendicular to the foot to avoid torsional misalignment.

The locking screw is inserted using the free-hand technique under fluoroscopy (**Fig. 56**).

Locate screw hole position under fluoroscopy and perform a 1-2cm incision. Bluntly dissect through the tissues to create a soft tissue tunnel to gain access to the bone. With the assistance of the tissue guide, drill bi-cortically ensuring the distal tip of the drill penetrates the far cortex.

Do not drill beyond the far cortex. Penetrate the far cortex only.

Measure the screw length using the gauge on the shaft of the drill.

Alternatively, screw length can be measured using the screw sizing gauge.

If the measurement reading sits between screw sizes, choose the longer one.

Using the short screwdriver with retention rod, insert the locking screw through the tissue guide and advance until the screw head is seated on the bone (**Fig. 56**).

Verify the locking screw position under fluoroscopy.

In instances of poor bone quality and subsequent insufficient fixation when using standard screws, the revision screw option can be utilized.

Remove the rotational reference half-pins.




Fig. 56 Free-hand screw positioning

Inserting the bone transport screw

The bone transport screw is inserted using a free-hand technique under fluoroscopy (**Fig. 57**). Please see page 30 for guidance - Inserting free-hand locking screw.




Fig. 57 Transport screw positioning

Inserting of additional free-hand locking screw

Additional locking screws are inserted using the same technique.

Fitbone Transport and Lengthening nail (TLN)

A maximum of two or three screws can be inserted, depending on the size.

Fitbone Transport nail (TN)

A maximum of three screws can be inserted.

Insert at least two screws to guarantee system stability (**Fig. 58**).




Fig. 58 Screw fixation

Jig removal

Part #	Description
60000689	Fitbone™ open end wrench SW14/17

Use the wrench to loosen the connection bolt on the jig (**Fig. 59**).




Fig. 59 Loosening the nail

Carefully remove the jig without damaging the cable (**Fig. 60**).




Fig. 60 Removing the jig

Cable management in suprapatellar approach

Make a 20mm transverse skin cut between lower border of patella and tuberosity ensuring preservation of the posterior cruciate ligament attachment (**Fig. 61**).

Split the patella tendon longitudinal for cable management.




Fig. 61 Additional skin cut in suprapatellar approach (L); patella tendon cut in suprapatellar approach (R)

Check AP and lateral where the cable is located and take it out with an arthroscopy hook (Fig. 62).



PRECAUTION: Do not hold the coupling or cable with surgical instruments and avoid bending the coupling or cable as this can lead to damage or unwanted disconnection.




Fig. 62 a) Cable position in suprapatellar approach; b) Splitting of the patella tendon; c) Detection of the cable.

Cable management in infrapatellar approach

For the infrapatellar approach the cable already exits from the patellar tendon, as shown in figure 63.

Securing the cable position at the patella tendon

Position the cable distally at the level of the tibial tuberosity.

Secure the cable in position by suturing the patella tendon together, directly above the cable (Fig. 63).



PRECAUTION: If the cable is not secured, patella tendon movement may result in damage to the cable




Fig. 63 Final cable position

Positioning the Receiver

Part #	Description
60001780	Fitbone subcutaneous energy receiver
60001622	Biotronik Torque Wrench 395121

Clean the plug of the cable properly in order to remove any blood (as it can act as an isolator).

Remove the transport locking device from the receiver's coupling. Place a drop of sterile water on the plug of the cable and insert into the coupling (Fig. 64).



PRECAUTION: Make sure the white ring (see blue circle in Fig. 64) is in connection with the start of the coupling.




Fig. 64 Transport locking device removal

Lock the cable into the Receiver's Coupling by tightening both screws with the torque wrench, until an audible click is heard (Fig. 65).



PRECAUTION: Do not hold the coupling or cable with surgical instruments and avoid bending the coupling or cable as this can lead to damage or unwanted disconnection.




Fig. 65 Receiver connection

The receiver should be placed antero-lateral directly underneath the skin. For this purpose, use scissors to prepare an 80 to 100mm subcutaneous skin pocket (Fig. 66).



PRECAUTION: Power transfer is optimal at approximately 5mm distance from the receiver. Avoid distances of more than 10mm under the skin as such distances can negatively affect the function of the treatment system.




Fig. 66 Subcutaneous skin pocket

Mark the position of the receiver on the skin (**Fig. 67**).




Fig. 67 Receiver final position

FIBULA OSTEOTOMY (only in bone transport + lengthening cases)

It is recommended that a screw (3.5-4mm short-threaded screw) for fixing the fibula to the tibia, is placed proximal to the syndesmosis in all additional lengthening cases (**Fig. 69**).

Perform a fibula osteotomy in the distal third, approximately 8cm above the joint line (**Fig. 68**). To ensure a complete osteotomy and avoid early fusion, it is recommended to resect approx. 5mm of fibula.




Fig. 68 Fibula osteotomy



Fig. 69 Fibula fixation

FINAL INTRAOPERATIVE TEST

Part #	Description
60001644	Fitbone Control Set for activation of intramedullary lengthening nail US plug 115V
60000676	Kawe stethoscope 06.10100.94

Upon skin closure perform an intraoperative functionality test as described above.

Place the transmitter and the stethoscope in separate sterile endoscopic camera drapes. **(Fig. 72 and 73)**
Perform a final functional test, ensuring the motor of the Fitbone can be heard during bone transport.

Control set settings for intraoperative test

Remove the switch cover cap from the control set. The surgeon can now choose between the patient/doctor mode respectively labelled "Pat." and "Doc." and pulse or continuous operation, respectively labelled "Pulse" and "Perm." **(Fig. 70)**

In the "Doc." position, the switch labelled "Doctor" on the front is enabled as well as the Pulse and Continuous modes. **(Fig. 71)**

Pressing the "Doctor" button while in the "Doc" and "Pulse" position will result in a continuous repetition of 1 second energy transmission and 9 sec pause between each distraction, until released.

Pressing the "Doctor" button while in the "Doc" and "Perm" position will result in a continuous energy transmission until released. This leads to a distraction rate of up to 2mm per minute. This mode can be used to preset the implant or to allow a new locking position for proximal locking. Use of continuous operation mode must be interrupted after a maximum of 1 minute for a minimum of 2 minutes to prevent excessive heat in the tissue between the Transmitter and Receiver. In doctor mode continuous operation, the Transmitter can reach a maximum temperature of 47.2°.

In the "Pat." position, the button labelled "Patient" on the front is enabled. In the "Pat." position, the permanent mode is blocked and will not lead to energy transmission.

Pressing the "Patient" button while in the "Pat." and "Pulse" position will result in a 90 second distraction cycle consisting of 9 repetitions of 1 second energy transmission and 9 sec break.

Pressing the "Patient" button while in the "Pat." and "Perm" position will not result in any energy transmission.

More information about the function and operation can be found in the instructions for use Fitbone Control Set.

For the intraoperative test, the doctor ("Doc.") and pulsating ("Pulse") settings are recommended. For the intraoperative final test, connect the transmitter to the Fitbone Control Set and place it on the skin directly above the receiver.



Fig. 70 Control Set settings



Fig. 71 Control Set



Fig. 72 Inserting the transmitter in the sterile endoscopic camera drape

Place the stethoscope on the patella. In the recommended settings, press the "Doctor" button and use the stethoscope to confirm correct functioning of the motor (**Fig. 74 and 75**).

Before handing over the Fitbone Control Set to the patient, set the switches to "Pat." and "Pulse" and put the switch cover back in place.

Thoroughly disinfect the Fitbone Control Set surface with a cloth moistened with 70% alcohol solution before handing over the set to the patient.



PRECAUTION: Please advise your patients not to remove the switch cover from the control set and not touch the switches.

Docking site procedure is optional, at the surgeon's discretion.



Fig. 73 Transmitter sterile covered



Fig 74 Doctor mode active conditions



Fig. 75 Final test under sterile

DOCKING SITE COMPRESSION AND HEALING

In accordance with the surgeon instructions the patient can continue bone transport to reach the docking site and compress both bone ends until cortical contact is achieved.

The compression protocols vary according to the patient needs and surgeon instruction. However, it is recommended to apply continuous compression until there is radiographical confirmation that cortical contact at the docking site has been achieved.

If needed 1mm compression can be applied until bone ends are touching. It is possible to repeat this procedure over days or weeks to ensure continuous bone to bone contact.

Compression and bone contact will restrict movement at the docking site. By the TN and TLN nails bone compression happens first at the end of excursion.

It is necessary to regularly monitor the docking site until union is achieved. Once the bone is consolidated and union at the docking site is achieved the nail can be exchanged for a trauma nail.




Fig. 76

LENGTHENING (OPTIONAL) WITH THE FITBONE TN/TLN

It is possible to perform bone lengthening with the Fitbone TLN nail variant.

Once the bone transport segment reaches the docking site, lengthening starts. Please refer to Fitbone Bone transport nail types for information on sizes and lengthening/transport possibilities.

“Rewind and Go”

Part #	Description
60001644	Fitbone Control Set for activation of intramedullary lengthening nail US plug 115V
60001871	FitboneControl Set for retraction of intramedullary lengthening nail
60000406	Fitbone™ screwdriver SW3.5 cannulated
60000576	Fitbone™ screwdriver SW3.5
60000384	Fitbone™ screw holder M2 L325
99-10013	Allen Wrench 3mm Sterile

If Rewind and Go procedure is needed as per pre-operative plan, the surgery is to be performed before bone consolidation is achieved, in order to prevent an unnecessary osteotomy.



Fig. 77

Rewind and Go refers to resetting the stroke by retraction to its starting position and repositioning the interlocking fixation at the bone segment so that the transport can be continued until docking.

The steps for *rewind and go* are:

1. External fixator placement (**78b proximal and 79a distal**). It is necessary to hold the transport segment in order to avoid the regenerate from recoiling.
2. Transport segment screw removal (**Fig. 78b and 79a**).
3. Nail retraction. (**Fig. 78c and 79b**) Please follow the instructions given at the "Retraction Control Set Quick Guide"
4. New screw insertion (**Fig. 78c or 79b**).




Fig. 78a Available transport stroke, 78b Proximal external fixator placement and screw removal, 78c Nail retraction and screw reinsertion



Fig. 79a Distal external fixator placement and screw removal, 79b Nail retraction and screw reinsertion

For example, a nail with an 80mm transport slot is selected; however, the size of the defect is 100mm. Initial transport can be carried out until the locking screw reaches the end of the transport slot (80mm). The patient is then brought back to the O.R. where the transport locking screw is removed, the motor retracted 20mm or more and a new locking screw is inserted. Transport is then carried out for another 20mm until the docking site is reached."

Please refer to "Quick Guide for Retraction Control Set" for further information."

The Retraction Control Set Components consist of the Control Electronics and a retraction transmitter. The Retraction Control Set Transmitter is larger and heavier than the transmitter of the control set (**Fig. 80**).



Fig. 80

Place the transmitter over the receiver so that they match symmetrically. The portion of the transmitter with the cable exiting should be in the same orientation and directly above the section of the receiver where the cable exits as shown. To ensure retraction, the orientation must be within $\pm 30^\circ$ of this position (green zone), otherwise distraction may occur (red zone) (**Fig. 81**).

Note: If the transmitter is angulated more than $\pm 30^\circ$ (red zone), nail distraction will take place. Failure to place the transmitter and the receiver concentrically will result in a reduced energy transfer and a decrease in retraction speed.




Fig. 81



PRECAUTION: If nail Rewind & Go is needed as per pre-operative plan, always replace the locking screw inserted in the bone transport segment with a new locking screw provided by the manufacturer.



PRECAUTION: During energy transmission, monitor the retraction with the stethoscope and continuous X-ray monitoring. It is also important to consider that locking screws inserted in the bone transport segment should be placed at least 10mm away from any newly formed cancellous bone.

FITBONE BONE TRANSPORT AND LENGTHENING SYSTEM – REMOVAL OPT – TIBIA

Patient positioning

Place the patient in supine position and cover the opposite side completely and the foot of the leg to be operated (Fig. 1).




Fig. 1 Patient positioned

Nail Marking

Using fluoroscopy, locate the nail and mark the position of the nail using a skin marker. The skin marking will aid orientation and positioning during surgery (Fig. 2).




Fig. 2 Nail marking tibia

Free-hand, transport and blocking screws removal

Part #	Description
700320	SHORT SCREWDRIVER
700319	SHORT SCREWDRIVER RETENTION ROD

Locate the screws under fluoroscopy and make a skin incision to gain access. If required, use a 5-8mm chisel to remove any bone obstructing access to the screws (Fig. 3).



Fig. 3 Free-hand and transport screws location and skin opening

Engage the first screw with the short screwdriver and retention rod and remove (**Fig. 4**). Repeat the same process to remove any remaining screws.




Fig. 4 Free-hand screw removal

Locate the transport screw under fluoroscopy and make a skin incision to gain access. Repeat the same process to remove it (**Fig. 5**).

Remove any blocking screws.




Fig. 5 Transport locking screw removal

Receiver removal

Locate and palpate the receiver on the skin and mobilize it in order to loosen any soft tissue that might be attached to it.

Make a 20mm skin incision along the scar of the receiver skin pocket (**Fig. 6**).




Fig. 6 Lateral cut

Locate and palpate along the cable to find the coupling (Fig. 7).




Fig. 7 Receiver location

Mobilize the coupling and remove from the skin pocket (Fig. 8). Use a clamp to secure the cable just before entering the patella tendon.




Fig. 8 Receiver extraction

Cut the cable proximally before the thicker part of the silicone (Fig. 9 and 10).

Remove the coupling and the receiver.




Fig. 9 Cable cut position




Fig. 10 Cutting the cable

Remove the clamp from the cable and secure a suture to the cable (**Fig. 11**).




Fig. 11 Securing the cable with a suture

Nail removal preparation - Infrapatellar approach - Leg positioning

Place the knee in maximum flexion during the surgical procedure (**Fig. 12**).




Fig. 12 Leg positioning infrapatellar approach

Approach

Split the patella tendon longitudinally (**Fig. 13**).




Fig. 13 Infrapatellar approach skin cut (left); Splitting of patella tendon (right)

Nail removal preparation - Suprapatellar approach - Leg positioning

In cases of patella baja supra patella approach might be required. The knee is flexed at about 30° during the surgical procedure (**Fig. 14**).




Fig. 14 Leg positioning suprapatellar approach

Approach

Make a 20mm transverse skin cut 20mm above the patella (**Fig. 15**).




Fig. 15 Suprapatellar approach skin cut

Split the quadriceps tendon longitudinally (**Fig. 16**).

Find the position of the nail inside the bone; if necessary, under fluoroscopy.




Fig. 16 Splitting of patella tendon

Connecting the nail and the Conical Mallet Adaptor

Part #	Description
183978	PLIER CUTTER D15MM
700336	M8 CONICAL MALLET ADAPTOR
91017	UNIVERSAL ALLEN WRENCH

Locate the position of the nail inside the bone. If required, use fluoroscopic guidance to assist identifying nail location.

Perform a blunt dissection to create a soft tissue tunnel to gain access to the top of the nail.

Use an arthroscopic hook to retrieve the cable/suture. Pass the suture through the cannulated reamer. Take up any suture slack to guide the reamer into position. Remove any newly formed bone or soft tissue blocking nail extraction pathway (**Fig. 16**).




Fig. 16

Pass the suture and cable through the Conical Mallet Adaptor (**Fig. 17**).




Fig. 17

Screw the Conical Mallet Adaptor clockwise into the nail and tighten it by using the universal Allen Wrench (**Fig. 18**).

Check with manual movements and under fluoroscopy, that the connection bolt is fully inserted into the nail. The connection bolt and nail should appear aligned longitudinally in the AP view.




Fig. 18

Nail extraction

Part #	Description
700320	SHORT SCREWDRIVER
700319	SHORT SCREWDRIVER RETENTION ROD
SMN 173370	SMN SYSTEM SLIDING HAMMER ASSEMBLY

Use the short screwdriver and retention rod to remove the first jig-assisted screw (**Fig. 19**).




Fig. 19 Cone positioning

Partially remove the final locking screw, ensuring the nail can pass upon extraction. Do not fully remove the screw, as should the connection to the nail become compromised, the screw can be reinserted to secure the nail once again. Connect the sliding hammer assembly to the conical mallet adaptor (**Fig. 20**).

If required, the broken screw, stripped thread screw and damaged screw head techniques described later in the operative technique can be used.

Use the sliding hammer to remove the nail. A combination of high repetitions and light to moderate force is advised.

Alternatively, it is possible to extract the nail with the slotted mallet by impacting the Conical Mallet Adapter.




Fig. 20 Working tube insertion

Upon nail extraction, fully remove the final locking screw (**Fig. 21**).

Flush the wound to remove any debris and confirm under fluoroscopy.

Perform a final check under fluoroscopy to ensure the site is free from debris.

Perform a layered skin closure of all surgical sites.





Fig. 21 Reaming of the bone

DAMAGED SCREW REMOVAL

Part #	Description
700320	SHORT SCREWDRIVER
700319	SHORT SCREWDRIVER RETENTION ROD
700501	HOLLOW REAMER D4.5MM
177380	SLOTTED MALLET
700503	SCREW HOLDER
183337	SCREW EXTRACTOR SIZE 3.5-9MM

Screw removal – broken screw



1. Shaft
2. centering pin
3. Reamer tube

Disengage the screw extraction sleeve (**Fig. 23**), remove the centering pin (component 2), and connect the shaft (components 1) and the reamer tube (component 3) to each other (**Fig. 24**).




Fig. 22 Broken screw




Fig. 23 Screw extraction sleeve disengagement




Fig. 24 Preparation of instrument components

Connect components 1 and 3 to the drill (**Fig. 25**).




Fig. 25 Connection to the drill bit

Carefully, approach the screw from the medial side. Place the hollow reamer over the tip of the screw and advance the reamer over the screw. Open the bone until reaching the broken screw (**Fig. 26**).

Disconnect the shaft (component 1) from the drill and leave the reamer tube (component 3) in place. Component 2 can now be inserted into component 3 and will move independently.




Fig. 26 Bone opening

Using a mallet and pliers, advance component 2 to push out the broken screw through the lateral side (**Fig. 27a,b and c**).




Fig. 27a, 27b, 27c Broken screw head removal

Screw removal – stripped thread

In the instance of a stripped thread screw, the screw holder (**Fig. 28**) can be used to engage and remove the screw.

With the screwdriver in place, advance the screw holder down the shaft of the driver and engage the screw head by sliding the teeth of the holder between the bone and screw head (**Fig. 29**).



Fig. 28 Screwdriver and screw holder




Figure 29 Stripped screw engagement

Remove the screw (**Fig. 30**).




Fig. 30 Stripped screw removal

Screw removal - damaged screw head

In the instance of a damaged screw head, use the screw extractor to engage the screw and remove (**Fig. 31**).

Rotate the screw extractor anti-clockwise to engage the screw head. Continue to rotate in the same direction to remove the screw.




Fig. 31 Screw extractor



FitboneTM

Intramedullary Lengthening System

Please refer to the **Instructions for Use PQFBT, PQFBP and PQFBR** supplied with the product for specific information on indications for use, contraindications, warnings, precautions, possible adverse events, MRI (Magnetic Resonance Imaging) safety information and sterilization.

Products may not be available in all markets because product availability is subject to regulatory and/or medical practices in individual markets. Please contact your Orthofix representatives if you have any questions about the availability of Orthofix products in your area.


Orthofix products or services referenced herein are trademarks or registered trademarks of Orthofix SRL and its group of companies.
All rights reserved.



* Manufactured by:
ORTHOFIX Srl
Via Delle Nazioni 9, 37012 Bussolengo
(Verona), Italy
Telephone: +39 045 6719000
Fax: +39 045 6719380
www.orthofix.com

* If not otherwise indicated on the product label

Distributed by:



Rx Only